

Dräger original accessories and consumables tested for compatibility

Dräger accessories & consumables are designed to be used with Dräger medical devices. In order to improve consistent operational performance and to reduce compatibility risk thus ensuring patient safety, Dräger not only tests the accessories which are developed to be used with the device, but also its consumable portfolio for compatibility with its medical devices.



A Declaration of Compatibility certifies that a certain product has been tested in combination with the corresponding device. The use of the device and the corresponding accessory is described in the Instruction for Use of both the device and the accessory.

That means for the operator of medical products – who has a particular focus on quality issues – that the correct functioning of original Dräger accessories & consumables has been tested for the respective device.

On this page you will find a selection of released consumables which have been tested with Dräger medical devices according to IEC 60601-2-13 and EN ISO 8835*.

DRÄGER QUALITY COMPONENTS

1. COMPATIBILITY
2. TESTCENTER
3. BIO-COMPATIBILITY
4. MATERIAL STRENGTH
5. MANUFACTURER SERVICE
6. USER TRAINING
7. SAFETY AT WORK
8. CERTIFICATES OF QUALITY

OXYLOG® 3000 PLUS

Consumables	Description	Order-No.
Breathing Circuits	VentStar® Oxylog 3000 Family & 2000 plus 150	5702871
	VentStar® Oxylog 3000 Family & 2000 plus 300	MP00335
	VentStar® Oxylog 3000 plus (P) 190	5704964
Filters/HMEs	SafeStar® 80	MP01785
	SafeStar® 55	MP01790
	SafeStar® 60A	MP01795
	TwinStar® 90	MP01800
	TwinStar® 55	MP01805
	TwinStar® 65A	MP01810
	TwinStar® 25	MP01815
	CareStar® 45	MP01755
	CareStar® 40A	MP01765
	CareStar® 30	MP01770



Oxylog® 3000 plus

*Please kindly note that the products listed on this page are only a selection. Additional products you will also find in the accessory list of the respective device. Products that are in the market for a long time and have been approved for use with the device before existence of IEC 60601-2-13 and EN ISO 8835 may not have a specific documented declaration of compatibility as they are part of the device specific accessory list. These accessories have already been evaluated during the CE Conformity Assessment procedure.

Masks	ComfortStar® Face Mask with Hook Ring, Size 1	MP01511
	ComfortStar® Face Mask with Hook Ring, Size 2	MP01512
	ComfortStar® Face Mask with Hook Ring, Size 3	MP01513
	ComfortStar® Face Mask with Hook Ring, Size 4	MP01514
	ComfortStar® Face Mask with Hook Ring, Size 5	MP01515
	ComfortStar® Face Mask with Hook Ring, Size 6	MP01516
	LiteStar® Anesthesia Face Mask without Hook Ring, Size 1	MP01501
	LiteStar® Anesthesia Face Mask without Hook Ring, Size 2	MP01502
	LiteStar® Anesthesia Face Mask without Hook Ring, Size 3	MP01503
	LiteStar® Anesthesia Face Mask without Hook Ring, Size 4	MP01504
	LiteStar® Anesthesia Face Mask without Hook Ring, Size 5	MP01505
	LiteStar® Anesthesia Face Mask without Hook Ring, Size 6	MP01506
	ClassicStar® NIV, Full-Face Mask with SE, Size S	MP01573
	ClassicStar® NIV, Full-Face Mask with SE, Size M	MP01574
	ClassicStar® NIV, Full-Face Mask with SE, Size L	MP01575
	NovaStar® NIV, Full-Face Mask with SE, Size S	MP01579
	NovaStar® NIV, Full-Face Mask with SE, Size M	MP01580
NovaStar® NIV, Full-Face Mask with SE, Size L	MP01581	
CO ₂ Cuvettes	Disposable CO ₂ Cuvette - Adult (Set of 10 pcs.)	MP01062
	Disposable CO ₂ Cuvette - Paediatric (Set of 10 pcs.)	MP01063

**HEADQUARTERS**

Dräger Medical AG & Co. KG
Moislinger Allee 53–55
23558 Lübeck, Germany

www.draeger.com

**REGION EUROPE CENTRAL
AND EUROPE NORTH**

Dräger Medical AG & Co. KG
Moislinger Allee 53–55
23558 Lübeck, Germany
Tel +49 451 882 0
Fax +49 451 882 2080
info@draeger.com

REGION EUROPE SOUTH

Dräger Médical S.A.S.
Parc de Haute
Technologie d'Antony 2
25, rue Georges Besse
92182 Antony Cedex
Tel +33 1 46 11 56 00
Fax +33 1 40 96 97 20
dlnfr-contact@draeger.com

**REGION MIDDLE EAST, AFRICA,
CENTRAL AND SOUTH AMERICA**

Dräger Medical AG & Co. KG
Dubai Healthcare City
P.O. Box 505108
Dubai, United Arab Emirates
Tel +971 436 24 762
Fax +971 436 24 761

REGION ASIA / PACIFIC

Dräger Medical
South East Asia Pte Ltd
25 International Business Park
#04-27/29 German Centre
Singapore 609916
Tel +65 6572 4388
Fax +65 6572 4399
asia.pacific@draeger.com

Manufacturer:

Dräger Medical AG & Co. KG
23542 Lübeck, Germany
The quality management system at Dräger Medical AG & Co. KG is certified according to ISO 13485, ISO 9001 and Annex II.3 of Directive 93/42/EEC (Medical devices).